## § 35.26

- (f) Licensees that are authorized for two or more different types of uses of byproduct material under Subparts E, F, and H of this part, or two or more types of units under Subpart H of this part, shall establish a Radiation Safety Committee to oversee all uses of byproduct material permitted by the license. The Committee must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service, and a representative of management who is neither an authorized user nor a Radiation Safety Officer. The Committee may include other members the licensee considers appro-
- (g) A licensee shall provide the Radiation Safety Officer sufficient authority, organizational freedom, time, resources, and management prerogative, to—
- (1) Identify radiation safety problems:
- (2) Initiate, recommend, or provide corrective actions;
  - (3) Stop unsafe operations; and,
- (4) Verify implementation of corrective actions.
- (h) A licensee shall retain a record of actions taken under paragraphs (a), (b), and (e) of this section in accordance with §35.2024.

## $\S\,35.26$ Radiation protection program changes.

- (a) A licensee may revise its radiation protection program without Commission approval if—
- (1) The revision does not require a license amendment under §35.13;
- (2) The revision is in compliance with the regulations and the license;
- (3) The revision has been reviewed and approved by the Radiation Safety Officer and licensee management; and
- (4) The affected individuals are instructed on the revised program before the changes are implemented.
- (b) A licensee shall retain a record of each change in accordance with §35.2026.

## §35.27 Supervision.

(a) A licensee that permits the receipt, possession, use, or transfer of byproduct material by an individual

- under the supervision of an authorized user, as allowed by \$35.11(b)(1), shall—
- (1) In addition to the requirements in §19.12 of this chapter, instruct the supervised individual in the licensee's written radiation protection procedures, written directive procedures, regulations of this chapter, and license conditions with respect to the use of byproduct material; and
- (2) Require the supervised individual to follow the instructions of the supervising authorized user for medical uses of byproduct material, written radiation protection procedures established by the licensee, written directive procedures, regulations of this chapter, and license conditions with respect to the medical use of byproduct material.
- (b) A licensee that permits the preparation of byproduct material for medical use by an individual under the supervision of an authorized nuclear pharmacist or physician who is an authorized user, as allowed by §35.11(b)(2), shall—
- (1) In addition to the requirements in §19.12 of this chapter, instruct the supervised individual in the preparation of byproduct material for medical use, as appropriate to that individual's involvement with byproduct material; and
- (2) Require the supervised individual to follow the instructions of the supervising authorized user or authorized nuclear pharmacist regarding the preparation of byproduct material for medical use, written radiation protection procedures established by the licensee, the regulations of this chapter, and license conditions.
- (c) A licensee that permits supervised activities under paragraphs (a) and (b) of this section is responsible for the acts and omissions of the supervised individual.

## §35.40 Written directives.

- (a) A written directive must be dated and signed by an authorized user before the administration of I-131 sodium iodide greater than 1.11 megabecquerels (MBq) (30 microcuries ( $\mu$ Ci)), any therapeutic dosage of unsealed byproduct material or any therapeutic dose of radiation from byproduct material.
- (1) If, because of the emergent nature of the patient's condition, a delay in